



Roche – Recall of Accu-Chek® Aviva Plus Test Strips

- On May 15, 2018, [Roche announced](#) a consumer level recall of several lots of [Accu-Chek Aviva Plus Test Strips](#) because the recalled strips had cracked reagent.
 - The cracked reagent in the affected strips may lead to an increased potential for strip errors prior to applying a blood sample, strips not detected once inserted, or in a very limited number of cases an inaccurate result may occur (which may not be detectable).
- The following products and lots are affected:

Product Description	Lot # (expiration date)
Accu-Chek Aviva Plus Strips	496947 (10/31/2018), 497291 (4/30/2019), 497325 (5/31/2019)
Accu-Chek Aviva Plus Health Network Strips	497296 (4/30/2019)

- Accu-Chek Aviva Plus test strips are used for blood glucose testing.
- If a defective strip is used for blood glucose testing and produces an inaccurate result, incorrect action could be taken by the patient.
- Patients should check the lot numbers of their test strip supply against the recalled lot numbers. The lot number can be found on the top flap of the vial packaging and on the label of each test strip vial.
- Patients should immediately discontinue use of and discard any recalled test strips. For replacement of recalled test strips, visit <https://notices.accu-chek.com> or email Accu-Chek Customer Care at accu-chek.care@roche.com.
- For more information regarding this recall, contact Accu-Chek Customer Care at **1-800-358-4866**.



OptumRx® specializes in the delivery, clinical management and affordability of prescription medications and consumer health products. We are an Optum® company — a leading provider of integrated health services. Learn more at optum.com.

All Optum® trademarks and logos are owned by Optum, Inc. All other brand or product names are trademarks or registered marks of their respective owners.

This document contains information that is considered proprietary to OptumRx and should not be reproduced without the express written consent of OptumRx.

RxNews® is published by the OptumRx Clinical Services Department.

©2018 Optum, Inc. All rights reserved.