



Simple Diagnostics – Recall of Alcohol Prep Pads

- On December 5, 2017, the [FDA announced](#) a voluntary, consumer-level recall of three lots of [Simple Diagnostics](#) Pharmacist Choice Alcohol Prep Pads, which were manufactured by Foshan Flying Medical Products, due to the lack of sterility assurance and other quality issues.
- The recalled lots were distributed between 10/18/2016 and 7/19/2017:

Product Name	NDC #	Lot # (Expiration Date)
Pharmacist Choice Alcohol Prep Pad	98302-0001-05	SD2070421201 (12/2019); SD2070420925 (9/2019); SD2070420601 (12/2019)

- Pharmacist Choice Alcohol Preps are supplied to distributors and pharmacies and used by healthcare professionals and patients for preparation of the skin prior to injection, as well as in first aid to decrease germs in minor cuts, scrapes and burns.
- The use of recalled Alcohol Prep Pads could result in adverse events such as infections. To date, Simple Diagnostics has not received any reports of adverse events associated with use of these alcohol prep pads.
- Anyone with recalled product should stop using the product and contact Simple Diagnostics for return information.
- For general questions regarding this recall and return information, contact Simple Diagnostics at **1-877-342-2385**.



OptumRx[®] specializes in the delivery, clinical management and affordability of prescription medications and consumer health products. We are an Optum[®] company — a leading provider of integrated health services. Learn more at [optum.com](#).

All Optum[®] trademarks and logos are owned by Optum, Inc. All other brand or product names are trademarks or registered marks of their respective owners.

This document contains information that is considered proprietary to OptumRx and should not be reproduced without the express written consent of OptumRx.

RxNews[®] is published by the OptumRx Clinical Services Department.

©2017 Optum, Inc. All rights reserved.