

## BD – Recall of BD Insulin Syringes with the BD Micro-Fine™ IV Needle

- On February 1, 2023, [BD announced](#) a consumer-level recall of some lots of BD Insulin Syringes with the BD Micro-Fine IV Needle because of manufacturing issues that can result in unsealed packaging for the individual syringes. If the blister pack of the syringe has a nonintact seal, then there is a potential that the syringe is no longer sterile.
  - Other BD insulin syringes that are not being recalled are available for patients to use.
- The recalled lots were distributed from June 2022 through October 2022.

Product Description	Lot# (Expiration Date)
BD Insulin Syringes with the BD Micro-Fine IV Needle 1 mL, 12.7 mm, 28 G	2038204 (4/30/27)
BD Insulin Syringes with the BD Micro-Fine IV Needle 1 mL, 12.7 mm, 28 G	2024074 (4/30/27)
BD Insulin Syringes with the BD Micro-Fine IV Needle 0.5 mL, 12.7 mm, 28 G	2031528 (4/30/27)

- Per BD, use of a non-sterile syringe poses the risk of a local skin infection which may require medical treatment such as the use of antibiotics.
- Anyone with an existing inventory of the recalled product should stop use, distribution and quarantine the product immediately.
- For any questions regarding this recall, contact BD by phone at **1-844-823-5433** or by email at [productcomplaints@bd.com](mailto:productcomplaints@bd.com).