



UltiMed – Withdrawal of UltiCare® pen needles

- On June 15, 2021, UltiMed announced a consumer-level withdrawal of one lot of UltiCare pen needles 6 mm x 31 G due to a report received that the inner box and case label expiration date is incorrectly labeled as February 28, 2006; the correct expiration date is February 28, 2026.
 - UltiCare pen needles that are not being withdrawn are available for patients to use.
- Affected product started shipping May 14, 2021.

Product Description	NDC#	Lot# (Expiration Date)
UltiCare Pen Needles 6 mm x 31 G Mini, 50 count	08222-0956-57	7210317 (2/28/2026)

- UltiCare pen needles are a diabetes care product used with diabetes pen delivery systems.
- Consumers should contact their physician or healthcare provider if they have experienced any problems that may be related to the withdrawn UltiCare pen needles.
- Anyone with an existing inventory of the withdrawn product should stop use, distribution and quarantine the product immediately.
- Contact UltiMed Customer Service by phone at **1-877-854-3434** or by email at **customerservice@ultimedinc.com** for more information about this withdrawal.



OptumRx® specializes in the delivery, clinical management and affordability of prescription medications and consumer health products. We are an Optum® company — a leading provider of integrated health services. Learn more at **optum.com**.

All Optum® trademarks and logos are owned by Optum, Inc. All other brand or product names are trademarks or registered marks of their respective owners.

This document contains information that is considered proprietary to OptumRx and should not be reproduced without the express written consent of OptumRx.

RxNews® is published by the OptumRx Clinical Services Department.

©2021 Optum, Inc. All rights reserved.