

B. Braun Medical – Recall of sodium chloride irrigation

- On February 16, 2018, [B. Braun Medical announced](#) a user-level recall of one lot of [sodium chloride irrigation](#) due to a customer complaint reporting particulate matter within the recalled lot. This particulate matter has been identified as polyethylene, which is consistent with the material used to manufacture the container cap.
- The recalled product was distributed November 3, 2017 through November 14, 2017.

| Product Description | NDC # | Lot # (expiration date) |
|---|--------------|----------------------------|
| 0.9% sodium chloride irrigation, 1000 mL plastic irrigation container | 0264-2201-00 | J7N912 (10/31/2020) |

- Sodium chloride irrigation is indicated for all general irrigation, washing, rinsing and dilution purposes which permit use of a sterile, nonpyrogenic electrolyte solution.
- Particulate matter present in the irrigation solution has the potential to cause a local inflammatory reaction. Particulate matter remaining in the tissue may cause the formation of a granuloma, with the potential for scar tissue formation and other local tissue insult.
- To date, B. Braun Medical has not received any reports of adverse events related to this recall.
- Anyone with recalled product should stop use and distribution and contact B. Braun Medical for return and replacement information.
- Contact B. Braun Medical at **1-800-227-2862** for more information regarding this recall.